

(10) Additional manufacturer narrative; and/or

(11) Corrected data, including:

(i) Any information missing on the user facility report or distributor report, including missing event codes, or information corrected on such forms after manufacturer verification;

(ii) For each event code provided by the user facility under § 803.32(d)(10) or a distributor, a statement of whether the type of the event represented by the code is addressed in the device labeling; and

(iii) If any required information was not provided, an explanation of why such information was not provided and the steps taken to obtain such information.

#### **§ 803.53 Five-day reports.**

A manufacturer shall submit a 5-day report to FDA, on Form 3500A or electronic equivalent as approved by FDA under § 803.14 within 5 workdays of:

(a) Becoming aware that a reportable MDR event or events, from any information, including any trend analysis, necessitates remedial action to prevent an unreasonable risk of substantial harm to the public health; or

(b) Becoming aware of an MDR reportable event for which FDA has made a written request for the submission of a 5-day report. When such a request is made, the manufacturer shall submit, without further requests, a 5-day report for all subsequent events of the same nature that involve substantially similar devices for the time period specified in the written request. The time period stated in the original written request can be extended by FDA if it is in the interest of the public health.

#### **§ 803.55 Baseline reports.**

(a) A manufacturer shall submit a baseline report on FDA Form 3417, or electronic equivalent as approved by FDA under § 803.14 for a device when the device model is first reported under § 803.50.

(b) Each baseline report shall be updated annually, on the anniversary month of the initial submission, after the initial baseline report is submitted. Changes to baseline information shall be reported in the manner described in

§ 803.56 (i.e., include only the new, changed, or corrected information in the appropriate portion(s) of the report form). Baseline reports shall contain the following:

(1) Name, complete address, and registration number of the manufacturer's reporting site. If the reporting site is not registered, FDA will assign a temporary registration number until the reporting site officially registers. The manufacturer will be informed of the temporary registration number;

(2) FDA registration number of each site where the device is manufactured;

(3) Name, complete address, and telephone number of the individual who has been designated by the manufacturer as its MDR contact and date of the report. For foreign manufacturers, a confirmation that the individual submitting the report is the agent of the manufacturer designated under § 803.58(a) is required;

(4) Product identification, including device family, brand name, generic name, model number, catalog number, product code and any other product identification number or designation;

(5) Identification of any device previously reported in a baseline report that is substantially similar (e.g., same device with a different model number, or same device except for cosmetic differences in color or shape) to the device being reported, including the identification of the previously reported device by model number, catalog number or other product identification, and the date of the baseline report for the previously reported device;

(6) Basis for marketing, including 510(k) premarket notification number or PMA number, if applicable, and whether the device is currently the subject of an approved post-market study under section 522 of the act;

(7) Date the device was initially marketed and, if applicable, the date on which the manufacturer ceased marketing the device;

(8) Shelf life, if applicable, and expected life of the device;

(9) The number of devices manufactured and distributed in the last 12 months and, an estimate of the number of devices in current use; and

(10) Brief description of any methods used to estimate the number of devices